




Analysis of Sucrose and Lactose in Chocolate Using HALO® 90 Å Penta-HILIC

216-SU



PEAK IDENTITIES:

1. Sucrose
2. D-(+) Lactose monohydrate

TEST CONDITIONS:

Column: HALO 90 Å Penta-HILIC, 2.7 μm , 4.6 x 150 mm

Part Number: 92814-705

Mobile Phase A: Water

Mobile Phase B: Acetonitrile

Flow Rate: 1.4 mL/min

Pressure: 213 bar

Temperature: 65 °C

Detection: ELSD, 40 °C, 3.3 bar

Injection Volume: 15 μL

Sample Solvent: 80/20 ACN/ Water

Response Time: 0.10 sec

DataRate: 10 Hz

LC System: Shimadzu NexeraX2

Chocolate is a very well-known, popular, food type worldwide. It is used for all occasions and can even have some health benefits as well, which include improved blood flow and brain function. There are four main types of chocolate to choose from- milk, white, dark, and raw.

Analysis of three different types of chocolate (milk, white, and dark) was carried out (or performed) in HILIC mode using an ELSD detector. The compounds of interest were sucrose and lactose. The HALO® Penta-HILIC column was used, which has a polar ligand with 5 hydroxyl groups tethered via novel proprietary linkage chemistry to Fused-Core® silica particles.

